

Created November 2016.

NAME

Alinity i HBsAg Qualitative II Controls (also referred to as HBsAg Qual Ctrls)

INTENDED USE

The Alinity i HBsAg Qualitative II Controls are for the estimation of test precision and the detection of systematic analytical deviations of the Alinity i analyzer when used for the qualitative detection and for the confirmation of the presence of hepatitis B surface antigen (HBsAg) in human serum and plasma using the Alinity i HBsAg Qualitative II and HBsAg Qualitative II Confirmatory Reagent Kits. For additional information, refer to the Alinity i HBsAg Qualitative II and Alinity i HBsAg Qualitative II Confirmatory reagent package inserts and the Alinity ci-series Operations Manual.

CONTENTS

CONTROL - contains recalcified human plasma. Preservatives: ProClin 950 and sodium azide.

CONTROL + contains inactivated purified human HBsAg (subtype *ad/ay*) in phosphate buffer with human plasma and protein (bovine serum albumin) stabilizers. Preservatives: ProClin 300 and ProClin 950.

The controls are at the following ranges:

Alinity i HBsAg Qualitative II (08P10)

Control	Quantity	Color	HBsAg	
			TARGET (S/CO)	RANGE (S/CO)
CONTROL -	1 x 8.0 mL	Natural	-	≤ 0.85
CONTROL +	1 x 8.0 mL	Blue ^a	3.50	1.75 - 5.25

^a Dye: Acid Blue No. 9

Alinity i HBsAg Qualitative II Confirmatory (08P11)

Control	Quantity	Color	HBsAg		%
			TARGET (S/CO ^b)	RANGE (S/CO ^b)	
CONTROL +	1 x 8.0 mL	Blue ^a	3.31	1.65 - 4.96	≥ 50%

^a Dye: Acid Blue No. 9

^b A target and a range are not defined for HBsAgQuaC1 S/CO.

NOTE: The insert ranges for the controls are not lot specific and represent the total range of values which may be generated throughout the life of the product. It is recommended that each laboratory establish its own means and acceptable ranges which should fall within the package insert ranges. Sources of variation that can be expected include:

- Calibration
- Control lot
- Reagent lot
- Calibrator lot
- Instrument

PRECAUTIONS

- **IVD**
- For *In Vitro* Diagnostic Use

Safety Precautions

-  **CAUTION:** This product contains human-sourced and/or potentially infectious components. Refer to the CONTENTS section of this package insert. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced materials should be considered potentially infectious. It is recommended that these reagents and human specimens be handled in accordance with the OSHA Standard on Bloodborne Pathogens. Biosafety Level 2 or other appropriate biosafety practices should be used for materials that contain or are suspected of containing infectious agents.¹⁻⁴
- The human-sourced material used in the Negative Control is nonreactive for HBsAg, HIV-1 Ag or HIV-1 RNA, anti-HIV-1/HIV-2, anti-HCV, and anti-HBs.
- The Positive Control contains purified HBsAg (inactivated) and human plasma. The human plasma is nonreactive for HBsAg, HIV-1 Ag or HIV-1 RNA, anti-HIV-1/HIV-2, anti-HCV, and anti-HBs.

The following warnings and precautions apply to: CONTROL -	
	
WARNING	Contains methylisothiazolone and sodium azide.
H317	May cause an allergic skin reaction.
EUH032	Contact with acids liberates very toxic gas.
Prevention	
P261	Avoid breathing mist / vapors / spray.
P272	Contaminated work clothing should not be allowed out of the workplace.
P280	Wear protective gloves / protective clothing / eye protection.
Response	
P302+P352	IF ON SKIN: Wash with plenty of water.
P333+P313	If skin irritation or rash occurs: Get medical advice / attention.
P362+P364	Take off contaminated clothing and wash it before reuse.
Disposal	
P501	Dispose of contents / container in accordance with local regulations.

The following warnings and precautions apply to: CONTROL +	
WARNING	Contains methylisothiazolones.
H317	May cause an allergic skin reaction.
Prevention	
P261	Avoid breathing mist / vapors / spray.
P272	Contaminated work clothing should not be allowed out of the workplace.
P280	Wear protective gloves / protective clothing / eye protection.
Response	
P302+P352	IF ON SKIN: Wash with plenty of water.
P333+P313	If skin irritation or rash occurs: Get medical advice / attention.
P362+P364	Take off contaminated clothing and wash it before reuse.
Disposal	
P501	Dispose of contents / container in accordance with local regulations.

Safety Data Sheets are available at www.abbottdiagnostics.com or contact your local representative.
For a detailed discussion of safety precautions during system operation, refer to the Alinity ci-series Operations Manual, Section 8.

PREPARATION FOR USE

- This product is liquid ready-to-use.
- This product may be used immediately after removal from 2 to 8°C storage.
- Prior to each use, mix by gentle inversion.

STORAGE

- Do not use past expiration date.

	Storage Temperature	Maximum Storage Time	Additional Storage Instructions
Unopened	2 to 8°C	Until expiration date	
Opened	2 to 8°C	Until expiration date	Store tightly capped. Store in an upright position. Return to refrigerated storage after use.

INSTRUMENT PROCEDURE

- To obtain the recommended volume requirements for the controls, hold the bottle vertically, and dispense 6 drops of the negative control and 6 drops of the positive control into each sample cup in the assigned position.
- For instructions on ordering and loading controls on the instrument, refer to the Alinity ci-series Operations Manual, Section 5.

INDICATIONS OF INSTABILITY OR DETERIORATION

Instability or deterioration should be suspected if there are precipitates, visible signs of leakage, turbidity, or if controls do not meet the appropriate package insert and/or Alinity ci-series Operations Manual criteria.

BIBLIOGRAPHY

1. US Department of Labor, Occupational Safety and Health Administration, 29 CFR Part 1910.1030, Bloodborne pathogens.
2. US Department of Health and Human Services. *Biosafety in Microbiological and Biomedical Laboratories*. 5th ed. Washington, DC: US Government Printing Office; December 2009.
3. World Health Organization. *Laboratory Biosafety Manual*. 3rd ed. Geneva: World Health Organization; 2004.
4. Clinical and Laboratory Standards Institute (CLSI). *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. CLSI Document M29-A4. Wayne, PA: CLSI; 2014.

Note for number formatting:

- A space is used as thousands separator (example: 10 000 specimens).
- A period is used to separate the integer part from the fractional part of a number written in decimal form (example: 3.12%).

Key to Symbols

	Caution
	Consult instructions for use
	Manufacturer
	Temperature limitation
	Use by/Expiration date
CN	Control Number
CONTAINS: AZIDE	Contains Sodium Azide. Contact with acids liberates very toxic gas.
CONTROL -	Negative Control
CONTROL +	Positive Control
IVD	<i>In Vitro</i> Diagnostic Medical Device
LOT	Lot Number
NEUTRALIZATION	Neutralization
PRODUCT OF IRELAND	Product of Ireland
RANGE	Range
REF	List Number
TARGET	Target

Alinity is a trademark of Abbott Laboratories in various jurisdictions. All other trademarks are property of their respective owners.

Abbott Ireland
Diagnostics Division
Finisklin Business Park
Sligo
Ireland
+353-71-9171712



Customer Service: Contact your local representative or find country-specific contact information on www.abbottdiagnostics.com

Created November 2016.

©2016 Abbott Laboratories