



MEMORANDUM	2026-093
Date:	15 October 2026
From:	Procurement Sector
To:	Bidders
Regarding:	Clarifications and Instructions to Bidders

Dear Bidders,

Please find below the clarifications and additional instructions regarding the tender:

1. Technical Clarifications

1. **Datasheets and Samples:** Only datasheets are required. No samples are needed.
2. **Item 11 – ECG Recorder:** The required device must support 12-lead ECG with simultaneous acquisition of all 12 leads, and a minimum 3-channel display and print. An additional option to connect the ECG to an external printer for report printing, in addition to printing on ECG paper, is preferable but not mandatory.
3. **Item 8 – Quantity:** The quantity indicated as zero is a typographical error. The correct quantity is **14**.
4. **CE/ISO Certification:** Please refer to Annex 3 for the applicable certification requirements.
5. **Medication Tablet Counter:** A manual tablet counter is required.
6. **Partial Offers:** Partial offers are accepted.

2. Instructions for Bid Submission

Please ensure that your bid complies with the following instructions:

1. Do not bind or staple any documents.
2. Clearly label all datasheets according to the corresponding item numbers.
3. Submit all documents required under Annex 7.
4. Do not modify the original LRC bidding template or add your company logo as a header.
5. Print all documents on one side of the paper only.
6. Submit your bid in a **sealed envelope to the Finance Sector** before the deadline specified in the tender documents.

Please ensure that all requirements and submission instructions are followed to avoid any issues during the evaluation process.

We wish you the best of luck.